

Manufacturer Disclosure Statement for Medical Device Security -- MDS2

Spacelabs Health

92877 091-0356-07 Rev A

May-24

Question ID	Question	See note	IEC TR 80001-2-2:2012	NIST SP 800-53 Rev. 4	ISO 27002:2013	
DOC-1	Manufacturer Name Device Description	Spacelabs Healthcare Intesys Clinical Suite (ICS) 12 Lead Electrocardiogram Version 5.6.1 (ECG) Interface is to provide a means to export 12-lead ECG reports and waveform data to an external system 12-lead database that is part of the hospital's information system.	—			
DOC-2	Device Model	92877	—			
DOC-3	Document ID	091-0356-07 Rev A	—			
DOC-4	Manufacturer Contact Information	Spacelabs Healthcare, 35301 SE Center Street, Snoqualmie, WA 98065 800-522-7025	—			
DOC-5	Intended use of device in network-connected environment:	12-Lead Electrocardiogram (ECG) Interface (92877) The intended use of the Spacelabs Healthcare 12-Lead Electrocardiogram (ECG) Interface is to provide a means to export 12-lead ECG reports and waveform data to an external system 12-lead database that is part of the hospital's information system.	—			
DOC-6	Document Release Date		—			
DOC-7	Coordinated Vulnerability Disclosure: Does the manufacturer have a vulnerability disclosure program for this device?		May-24			
DOC-8	ISAO: Is the manufacturer part of an Information Sharing and Analysis Organization?	Yes				
DOC-9	Diagram: Is a network or data flow diagram available that indicates connections to other system components or expected external resources?	No				
DOC-10	SaMD: Is the device Software as a Medical Device (i.e. software-only, no hardware)?	Yes				
DOC-11	Does the SaMD contain an operating system?	No				
DOC-11.1	Does the SaMD rely on an owner/operator provided operating system?					
DOC-11.2	Is the SaMD hosted by the manufacturer?	Yes				
DOC-11.3	Is the SaMD hosted by the customer?	No				
DOC-11.4	Is the SaMD hosted by the manufacturer?	Yes				
		Yes, No, N/A, or See Note	Note #			
	MANAGEMENT OF PERSONALLY IDENTIFIABLE INFORMATION					
	Can this device display, transmit, store, or modify personally identifiable information (e.g. electronic Protected Health Information (ePHI))?					
MPII-1	Does the device maintain personally identifiable information?	Yes				
MPII-2	Does the device maintain personally identifiable information temporarily in volatile memory (i.e., until cleared by power-off or reset)?	Yes				
MPII-2.1	Does the device store personally identifiable information persistently on internal media?	Yes				
MPII-2.2		Yes				
				IEC TR 80001-2-2:2012	NIST SP 800-53 Rev. 4	ISO 27002:2013
					AR-2	A.15.1.4
					AR-2	A.15.1.4
					AR-2	A.15.1.4

MPII-2.3	Is personally identifiable information preserved in the device's non-volatile memory until explicitly	No	—		
MPII-2.4	Does the device store personally identifiable information in a database?	Yes	—		
MPII-2.5	Does the device allow configuration to automatically delete local personally identifiable information after it is stored to a long term solution?	No	—	AR-2	A.15.1.4
MPII-2.6	Does the device import/export personally identifiable information with other systems (e.g., a wearable monitoring device might export personally identifiable information to a server)?	Yes	—	AR-2	A.15.1.4
MPII-2.7	Does the device maintain personally identifiable information when powered off, or during power service interruptions?	No	—	AR-2	A.15.1.4
MPII-2.8	Does the device allow the internal media to be removed by a service technician (e.g., for separate destruction or customer retention)?	N/A	—		
MPII-2.9	Does the device allow personally identifiable information records be stored in a separate location from the device's operating system (i.e. secondary internal drive, alternate drive partition, or remote storage location)?	No	The information is stored in the database, which is a different component within ICS.	AR-2	A.15.1.4
MPII-3	Does the device have mechanisms used for the transmitting, importing/exporting of personally identifiable information?	Yes	—	AR-2	A.15.1.4
MPII-3.1	Does the device display personally identifiable information (e.g., video display, etc.)?	No	—	AR-2	A.15.1.4
MPII-3.2	Does the device generate hardcopy reports or images containing personally identifiable information?	No	—	AR-2	A.15.1.4
MPII-3.3	Does the device retrieve personally identifiable information from or record personally identifiable information to removable media (e.g., removable-HDD, USB memory, DVD-R/RW, CD-R/RW, tape, CF/SD card, memory stick, etc.)?	No	—	AR-2	A.15.1.4
MPII-3.4	Does the device transmit/receive or import/export personally identifiable information via dedicated cable connection (e.g., RS-232, RS-423, USB, etc.)?	No	—	AR-2	A.15.1.4
MPII-3.5	Does the device transmit/receive personally identifiable information via a wired network connection (e.g., RJ45, fiber optic, etc.)?	Yes	—	AR-2	A.15.1.4
MPII-3.6	Does the device transmit/receive personally identifiable information via a wireless network connection (e.g., WiFi, Bluetooth, NFC, infrared, cellular, etc.)?	No	—	AR-2	A.15.1.4
MPII-3.7	Does the device transmit/receive personally identifiable information over an external network (e.g., Internet)?	No	ICS receives data from wired or wireless monitoring devices over the network deployed onsite	AR-2	A.15.1.4
MPII-3.8	Does the device import personally identifiable information via scanning a document?	No			
MPII-3.9	Does the device transmit/receive personally identifiable information via a proprietary protocol?	No			
MPII-3.10	Does the device use any other mechanism to transmit, import or export personally identifiable information?	No	—	AR-2	A.15.1.4
Management of Private Data notes:				AR-2	A.15.1.4

AUTOMATIC LOGOFF (ALOF)

The device's ability to prevent access and misuse by unauthorized users if device is left idle for a period of time.

Can the device be configured to force reauthorization of logged-in user(s) after a predetermined length of inactivity (e.g., auto-logoff, session lock, password protected screen saver)?

ALOF-1	No	—		Section 5.1, ALOF	AC-12	None
--------	----	---	--	-------------------	-------	------

IEC TR 80001-2-2:2012

NIST SP 800-53 Rev. 4

ISO 27002:2013

ALOF-2	Is the length of inactivity time before auto-logout/screen lock user or administrator	N/A	—	Section 5.1, ALOF	AC-11	A.11.2.8, A.11.2.9
AUDIT CONTROLS (AUDT)				IEC TR 80001-2-2:2012	NIST SP 800-53 Rev. 4	ISO 27002:2013
<i>The ability to reliably audit activity on the device.</i>						
AUDT-1	Can the medical device create additional audit logs or reports beyond standard operating system logs?	Yes	—	Section 5.2, AUDT	AU-1	A.5.1.1, A.5.1.2, A.6.1.1, A.12.1.1, A.18.1.1, A.18.2.2
AUDT-1.1	Does the audit log record a USER ID?	Yes	—			
AUDT-1.2	Does other personally identifiable information exist in the audit trail?	No		Section 5.2, AUDT	AU-2	None
	Are events recorded in an audit log? If yes, indicate which of the following events are recorded in the audit log:	Yes	—	Section 5.2, AUDT	AU-2	None
AUDT-2	Successful login/logout attempts?	Yes	—	Section 5.2, AUDT	AU-2	None
AUDT-2.1	Unsuccessful login/logout attempts?	No	—	Section 5.2, AUDT	AU-2	None
AUDT-2.2	Modification of user privileges?	No	—	Section 5.2, AUDT	AU-2	None
AUDT-2.3	Creation/modification/deletion of users?	No	—	Section 5.2, AUDT	AU-2	None
AUDT-2.4	Presentation of clinical or PII data (e.g. display, Creation/modification/deletion of data)?	No	—	Section 5.2, AUDT	AU-2	None
AUDT-2.5	Import/export of data from removable media (e.g. USB drive, external hard drive, DVD)?	N/A	—	Section 5.2, AUDT	AU-2	None
AUDT-2.6	Receipt/transmission of data or commands over a network or point-to-point connection?	Yes		Section 5.2, AUDT	AU-2	None
AUDT-2.7	Remote or on-site support?		ICS runs on COTS software. Remote Access to the various ICS servers can be configured by the customer numerous ways. Many of these provide detailed logging of remote access. This is not a service ICS provides	Section 5.2, AUDT	AU-2	None
AUDT-2.8	Application Programming Interface (API) and similar activity?	No				
AUDT-2.8.1	Emergency access?	No	—	Section 5.2, AUDT	AU-2	None
AUDT-2.8.2	Other events (e.g., software updates)?	N/A	—	Section 5.2, AUDT	AU-2	None
AUDT-2.9	Is the audit capability documented in more detail?	No	—	Section 5.2, AUDT	AU-2	None
AUDT-2.10	Can the owner/operator define or select which events are recorded in the audit log?	No		Section 5.2, AUDT	AU-2	None
AUDT-2.11	Is a list of data attributes that are captured in the audit log for an event available?	No	—	Section 5.2, AUDT	AU-2	None
AUDT-3	Does the audit log record date/time?	Yes	—	Section 5.2, AUDT	AU-2	None
AUDT-4	Can date and time be synchronized by Network Time Protocol (NTP) or equivalent time source?	Yes	—	Section 5.2, AUDT	AU-2	None
AUDT-4.1	Can audit log content be exported?	No	—	Section 5.2, AUDT	AU-2	None
AUDT-4.1.1	Can audit log content be exported?	No	—	Section 5.2, AUDT	AU-2	None
AUDT-5	Can audit log content be exported?	No	—	Section 5.2, AUDT	AU-2	None
AUDT-5.1	Via physical media?	No	—			
AUDT-5.2	Via IHE Audit Trail and Node Authentication (ATNA) profile to SIEM?	No	—			
AUDT-5.3	Via Other communications (e.g., external service device, mobile applications)?	No	—			
AUDT-5.4	Are audit logs encrypted in transit or on storage media?	No	—			
AUDT-6	Can audit logs be monitored/reviewed by owner/operator?	No	—			
AUDT-7	Are audit logs protected from modification?	N/A	—	Section 5.2, AUDT	AU-2	None
AUDT-7.1	Are audit logs protected from access?	Yes	—			
AUDT-8	Can audit logs be analyzed by the device?	No	—	Section 5.2, AUDT	AU-2	None
AUTHORIZATION (AUTH)				IEC TR 80001-2-2:2012	NIST SP 800-53 Rev. 4	ISO 27002:2013
<i>The ability of the device to determine the authorization of users.</i>						
AUTH-1	Does the device prevent access to unauthorized users through user login requirements or other mechanism?	No	—	Section 5.3, AUTH	IA-2	A.9.2.1

AUTH-1.1	Can the device be configured to use federated credentials management of users for authorization (e.g., LDAP, OAuth)?	Yes	—	Section 5.3, AUTH	IA-2	A.9.2.1
AUTH-1.2	Can the customer push group policies to the device (e.g., Active Directory)?	Yes	—	Section 5.3, AUTH	IA-2	A.9.2.1
AUTH-1.3	Are any special groups, organizational units, or group policies required?	No	—	Section 5.3, AUTH	IA-2	A.9.2.1
AUTH-2	Can users be assigned different privilege levels based on 'role' (e.g., user, administrator, and/or service, etc.)?	No	—	Section 5.3, AUTH	IA-2	A.9.2.1
AUTH-3	Can the device owner/operator grant themselves unrestricted administrative privileges (e.g., access operating system or application via local root or administrator account)?	Yes	—	Section 5.3, AUTH	IA-2	A.9.2.1
AUTH-4	Does the device authorize or control all API access requests?	No	—	Section 5.3, AUTH	IA-2	A.9.2.1
AUTH-5	Does the device run in a restricted access mode, or 'kiosk mode', by default?	No	—			

CYBER SECURITY PRODUCT UPGRADES (CSUP)

The ability of on-site service staff, remote service staff, or authorized customer staff to install/upgrade device's security patches.

IEC TR 80001-2-2:2012

NIST SP 800-53 Rev. 4

ISO 27002:2013

CSUP-1	Does the device contain any software or firmware which may require security updates during its operational life, either from the device manufacturer or from a third-party manufacturer of the software/firmware? If no, answer "N/A" to questions in this section.	Yes	—			
CSUP-2	Does the device contain an Operating System? If yes, complete 2.1-2.4.	Yes	Operating system requirement is available in the product datasheet			
CSUP-2.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	Yes	—			
CSUP-2.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	Yes	—			
CSUP-2.3	Does the device have the capability to receive remote installation of patches or software updates?	No	—			
CSUP-2.4	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	Yes	Spacelabs conducts monthly Microsoft patch verification testing for our Windows-based products. It is recommended to review the patch test reports prior to patching in the event an update can cause impact to the hosted Spacelabs product. The patch test reports can be found here - https://www.spacelabshealthcare.com/products/security/patch-test-reports-access-form/?redirect_to=%2Fproducts%2Fsecurity%2Fpatch-test-reports%2F			
CSUP-3	Does the device contain Drivers and Firmware? If yes, complete 3.1-3.4.	No	—			
CSUP-3.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	N/A	—			
CSUP-3.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	No	Product software is hosted at customer. Updates are managed by the manufacturer or by an authorized representative. Windows Operating System updates are managed by the customer.			
CSUP-3.3	Does the device have the capability to receive remote installation of patches or software updates?	No	—			
CSUP-3.4	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	Yes	—			

CSUP-4	Does the device contain Anti-Malware Software? If yes, complete 4.1-4.4.	No	Product host server can have any commercial anti-malware product installed. Please see the ICS Anti-malware customer service notice for recommended client configuration.			
CSUP-4.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	No	Product software updates are managed by the manufacturer or by an authorized representative.			
CSUP-4.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	No	—			
CSUP-4.3	Does the device have the capability to receive remote installation of patches or software updates?	No	—			
CSUP-4.4	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	Yes	—			
CSUP-5	Does the device contain Non-Operating System commercial off-the-shelf components? If yes, complete 5.1-5.4.	Yes	—			
CSUP-5.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	No	—			
CSUP-5.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	Yes	—			
CSUP-5.3	Does the device have the capability to receive remote installation of patches or software updates?	No	—			
CSUP-5.4	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	Yes	—			
CSUP-6	Does the device contain other software components (e.g., asset management software, license management)? If yes, please provide details or reference in notes and complete 6.1-6.4.	No	—			
CSUP-6.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	N/A	—			
CSUP-6.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	N/A	—			
CSUP-6.3	Does the device have the capability to receive remote installation of patches or software updates?	N/A	—			
CSUP-6.4	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	N/A	—			
CSUP-7	Does the manufacturer notify the customer when updates are approved for installation?	Yes	Spacelabs conducts monthly Microsoft patch verification testing for our Windows-based products. It is recommended to review the patch test reports prior to patching in the event an update can cause impact to the hosted Spacelabs product. The patch test reports can be found here - https://www.spacelabshealthcare.com/products/security/patch-test-reports-access-form/?redirect_to=%2Fproducts%2Fsecurity%2Fpatch-test-reports%2F			
CSUP-8	Does the device perform automatic installation of software updates?	No	—			
CSUP-9	Does the manufacturer have an approved list of third-party software that can be installed on the device?	N/A	—			
CSUP-10	Can the owner/operator install manufacturer-approved third-party software on the device themselves?	N/A	—			
CSUP-10.1	Does the system have mechanism in place to prevent installation of unapproved software?	N/A	—			
CSUP-11	Does the manufacturer have a process in place to assess device vulnerabilities and updates?	Yes	—			

CSUP-11.1	Does the manufacturer provide customers with review and approval status of updates?	No	Spacelabs conducts monthly Microsoft patch verification testing for our Windows-based products. It is recommended to review the patch test reports prior to patching in the event an update can cause impact to the hosted Spacelabs product. The patch test reports can be found here - https://www.spacelabshealthcare.com/products/security/patch-test-reports-access-form/?redirect_to=%2Fproducts%2Fsecurity%2Fpatch-test-reports%2F			
CSUP-11.2	Is there an update review cycle for the device?	No				

HEALTH DATA DE-IDENTIFICATION (DIDT)

The ability of the device to directly remove information that allows identification of a person.

IEC TR 80001-2-2:2012

NIST SP 800-53 Rev. 4

ISO 27002:2013

DIDT-1	Does the device provide an integral capability to de-identify personally identifiable information?	No	—	Section 5.6, DIDT	None	ISO 27038
DIDT-1.1	Does the device support de-identification profiles that comply with the DICOM standard for de-identification?	No	—	Section 5.6, DIDT	None	ISO 27038

DATA BACKUP AND DISASTER RECOVERY (DTBK)

The ability to recover after damage or destruction of device data, hardware, software, or site configuration information.

IEC TR 80001-2-2:2012

NIST SP 800-53 Rev. 4

ISO 27002:2013

DTBK-1	Does the device maintain long term primary storage of personally identifiable information / patient information (e.g. PACS)?	No	—			
DTBK-2	Does the device have a "factory reset" function to restore the original device settings as provided by the manufacturer?	No	—	Section 5.7, DTBK	CP-9	A.12.3.1
DTBK-3	Does the device have an integral data backup capability to removable media?	No	—	Section 5.7, DTBK	CP-9	A.12.3.1
DTBK-4	Does the device have an integral data backup capability to remote storage?	No	—			
DTBK-5	Does the device have a backup capability for system configuration information, patch restoration, and software restoration?	No	—			
DTBK-6	Does the device provide the capability to check the integrity and authenticity of a backup?	No	—	Section 5.7, DTBK	CP-9	A.12.3.1

EMERGENCY ACCESS (EMRG)

The ability of the device user to access personally identifiable information in case of a medical emergency situation that requires immediate access to stored personally identifiable information.

IEC TR 80001-2-2:2012

NIST SP 800-53 Rev. 4

ISO 27002:2013

EMRG-1	Does the device incorporate an emergency access (i.e. "break-glass") feature?	No	—	Section 5.8, EMRG	SI-17	None
--------	---	----	---	-------------------	-------	------

HEALTH DATA INTEGRITY AND AUTHENTICITY (IGAU)

How the device ensures that the stored data on the device has not been altered or destroyed in a non-authorized manner and is from the originator.

IEC TR 80001-2-2:2012

NIST SP 800-53 Rev. 4

ISO 27002:2013

IGAU-1	Does the device provide data integrity checking mechanisms of stored health data (e.g., hash or digital signature)?	No	—	Section 5.9, IGAU	SC-28	A.18.1.3
--------	---	----	---	-------------------	-------	----------

IGAU-2	Does the device provide error/failure protection and recovery mechanisms for stored health data (e.g., RAID-5)?	No	—	Section 5.9, IGAU	SC-28	A.18.1.3
MALWARE DETECTION/PROTECTION (MLDP) <i>The ability of the device to effectively prevent, detect and remove malicious software (malware).</i>				IEC TR 80001-2-2:2012	NIST SP 800-53 Rev. 4	ISO 27002:2013
MLDP-1	Is the device capable of hosting executable software?	Yes	—	Section 5.10, MLDP		
	Does the device support the use of anti-malware software (or other anti-malware mechanism)? Provide details or reference in notes.		Product host server can have any commercial anti-malware product installed. Please see the ICS Anti-malware customer service notice for recommended client configuration.			
MLDP-2	Does the device include anti-malware software by default?	Yes		Section 5.10, MLDP	SI-3	A.12.2.1
MLDP-2.1	Does the device have anti-malware software available as an option?	No	—	Section 5.10, MLDP	CM-5	A.9.2.3, A.9.4.5, A.12.1.2, A.12.1.4, A.12.5.1
MLDP-2.2	Does the device documentation allow the owner/operator to install or update anti-malware software?	N/A	—	Section 5.10, MLDP	AU-6	A.12.4.1, A.16.1.2, A.16.1.4
MLDP-2.3	Can the device owner/operator independently (re-)configure anti-malware settings?	N/A	—	Section 5.10, MLDP	CP-10	A.17.1.2
MLDP-2.4	Does notification of malware detection occur in the device user interface?	N/A	—	Section 5.10, MLDP	AU-2	None
MLDP-2.5	Can only manufacturer-authorized persons repair systems when malware has been detected?	No				
MLDP-2.6	Are malware notifications written to a log?	N/A				
MLDP-2.7	Are there any restrictions on anti-malware (e.g., purchase, installation, configuration, scheduling)?		Product host server can have any commercial anti-malware product installed. Please see the ICS Anti-malware customer service notice for recommended client configuration.			
MLDP-2.8	If the answer to MLDP-2 is NO, and anti-malware cannot be installed on the device, are other compensating controls in place or available?	Yes				A.12.6.1, A.14.2.2, A.14.2.3, A.16.1.3
MLDP-3	Does the device employ application whitelisting that restricts the software and services that are permitted to be run on the device?	N/A	—	Section 5.10, MLDP	SI-2	
MLDP-4	Does the device employ a host-based intrusion detection/prevention system?	N/A	—	Section 5.10, MLDP	SI-3	A.12.2.1
MLDP-5	Can the host-based intrusion detection/prevention system be configured by the customer?	N/A	—	Section 5.10, MLDP	SI-4	None
MLDP-5.1	Can a host-based intrusion detection/prevention system be installed by the customer?	N/A	—	Section 5.10, MLDP	CM-7	A.12.5.1
MLDP-5.2		N/A	—	Section 5.10, MLDP		
NODE AUTHENTICATION (NAUT) <i>The ability of the device to authenticate communication partners/nodes.</i>				IEC TR 80001-2-2:2012	NIST SP 800-53 Rev. 4	ISO 27002:2013
NAUT-1	Does the device provide/support any means of node authentication that assures both the sender and the recipient of data are known to each other and are authorized to receive transferred information (e.g. Web APIs, SMTP, SNMP)?	No	—	Section 5.11, NAUT	SC-23	None
NAUT-2	Are network access control mechanisms supported (E.g., does the device have an internal firewall, or use a network connection white list)?	N/A	—	Section 5.11, NAUT	SC-7	A.13.1.1, A.13.1.3, A.13.2.1, A.14.1.3
NAUT-2.1	Is the firewall ruleset documented and available for review?	N/A	—			
NAUT-3	Does the device use certificate-based network connection authentication?	No	—			
CONNECTIVITY CAPABILITIES (CONN)				IEC TR 80001-2-2:2012	NIST SP 800-53 Rev. 4	ISO 27002:2013

All network and removable media connections must be considered in determining appropriate security controls. This section lists connectivity capabilities that may be present on the device.

CONN-1	Does the device have hardware connectivity capabilities?	No	ICS is a software product.
CONN-1.1	Does the device support wireless connections?	N/A	—
CONN-1.1.1	Does the device support Wi-Fi?	N/A	—
CONN-1.1.2	Does the device support Bluetooth?	N/A	—
CONN-1.1.3	Does the device support other wireless network connectivity (e.g. LTE, Zigbee, proprietary)?	N/A	—
CONN-1.1.4	Does the device support other wireless connections (e.g., custom RF controls, wireless detectors)?	N/A	—
CONN-1.2	Does the device support physical connections?	N/A	—
CONN-1.2.1	Does the device have available RJ45 Ethernet ports?	No	—
CONN-1.2.2	Does the device have available USB ports?	N/A	—
CONN-1.2.3	Does the device require, use, or support removable memory devices?	N/A	—
CONN-1.2.4	Does the device support other physical connectivity?	N/A	—
	Does the manufacturer provide a list of network ports and protocols that are used or may be used on the device?		ICS is a software product that will be hosted on customer hardware. Spacelabs can provide the necessary ports and protocols for customers to configure.
CONN-2	Can the device communicate with other systems within the customer environment?	N/A	ICS is a software product that will be hosted on customer hardware.
CONN-3	Can the device communicate with other systems external to the customer environment (e.g., a service host)?	No	—
CONN-4	Does the device make or receive API calls?	Yes	—
CONN-5	Does the device require an internet connection for its intended use?	No	—
CONN-6	Does the device support Transport Layer Security (TLS)?	Yes	—
CONN-7.1	Is TLS configurable?	N/A	—
CONN-8	Does the device provide operator control functionality from a separate device (e.g., telemedicine)?	No	—

PERSON AUTHENTICATION (PAUT)

The ability to configure the device to authenticate users.

PAUT-1	Does the device support and enforce unique IDs and passwords for all users and roles (including service accounts)?	Yes	—
PAUT-1.1	Does the device enforce authentication of unique IDs and passwords for all users and roles (including service accounts)?	Yes	—
PAUT-2	Is the device configurable to authenticate users through an external authentication service (e.g., MS Active Directory, NDS, LDAP, OAuth, etc.)?	Yes	—
PAUT-3	Is the device configurable to lock out a user after a certain number of unsuccessful logon attempts?	No	—
PAUT-4	Are all default accounts (e.g., technician service accounts, administrator accounts) listed in the documentation?	N/A	—
PAUT-5	Can all passwords be changed?	N/A	—
PAUT-6	Is the device configurable to enforce creation of user account passwords that meet established (organization specific) complexity rules?	N/A	—
PAUT-7	Does the device support account passwords that expire periodically?	N/A	—
PAUT-8	Does the device support multi-factor authentication?	N/A	—
PAUT-9	Does the device support single sign-on (SSO)?	Yes	—
PAUT-10	Can user accounts be disabled/locked on the device?	No	—

IEC TR 80001-2-2:2012

NIST SP 800-53 Rev. 4

ISO 27002:2013

Section 5.12, PAUT

IA-2

A.9.2.1

Section 5.12, PAUT

IA-2

A.9.2.1

Section 5.12, PAUT

IA-5

A.9.2.1

Section 5.12, PAUT

IA-2

A.9.2.1

Section 5.12, PAUT

SA-4(5)

A.14.1.1, A.14.2.7, A.14.2.9, A.15.1.2

Section 5.12, PAUT

IA-2

A.9.2.1

Section 5.12, PAUT

IA-2

A.9.2.1

Section 5.12, PAUT

IA-2

A.9.2.1

Spacelabs Health 92877 091-0356-07 Rev A May-24

PAUT-11	Does the device support biometric controls?	N/A	—
PAUT-12	Does the device support physical tokens (e.g. badge access)?	No	—
PAUT-13	Does the device support group authentication (e.g. hospital teams)?	No	—
PAUT-14	Does the application or device store or manage authentication credentials?	No	—
PAUT-14.1	Are credentials stored using a secure method?	N/A	—

Section 5.12, PAUT IA-2 A.9.2.1

PHYSICAL LOCKS (PLOK)

Physical locks can prevent unauthorized users with physical access to the device from compromising the integrity and confidentiality of personally identifiable information stored on the device or on removable media

IEC TR 80001-2-2:2012 NIST SP 800-53 Rev. 4 ISO 27002:2013

PLOK-1	Is the device software only? If yes, answer “N/A” to remaining questions in this section. Are all device components maintaining personally identifiable information (other than removable media) physically secure (i.e., cannot remove without tools)?	Yes	—
PLOK-2	Are all device components maintaining personally identifiable information (other than removable media) physically secured behind an individually keyed locking device?	N/A	—
PLOK-3	Does the device have an option for the customer to attach a physical lock to restrict access to removable media?	N/A	—
PLOK-4		N/A	—

Section 5.13, PLOK PE- 3(4) A.11.1.1, A.11.1.2, A.11.1.3
 Section 5.13, PLOK PE- 3(4) A.11.1.1, A.11.1.2, A.11.1.3
 Section 5.13, PLOK PE- 3(4) A.11.1.1, A.11.1.2, A.11.1.3
 Section 5.13, PLOK PE- 3(4) A.11.1.1, A.11.1.2, A.11.1.3

ROADMAP FOR THIRD PARTY COMPONENTS IN DEVICE LIFE CYCLE (RDMP)

Manufacturer’s plans for security support of third-party components within the device’s life cycle.

Was a secure software development process, such as ISO/IEC 27034 or IEC 62304, followed during product development?

IEC TR 80001-2-2:2012 NIST SP 800-53 Rev. 4 ISO 27002:2013

RDMP-1	Does the manufacturer evaluate third-party applications and software components included in the device for secure development practices?	Yes	—
RDMP-2	Does the manufacturer maintain a web page or other source of information on software support dates and updates?	Yes	—
RDMP-3	Does the manufacturer have a plan for managing third-party component end-of-life?	N/A	—
RDMP-4		N/A	—

Section 5.14, RDMP CM-2 None
 Section 5.14, RDMP CM-8 A.8.1.1, A.8.1.2
 Section 5.14, RDMP CM-8 A.8.1.1, A.8.1.2
 Section 5.14, RDMP CM-8 A.8.1.1, A.8.1.2

SOFTWARE BILL OF MATERIALS (SBoM)

A Software Bill of Material (SBoM) lists all the software components that are incorporated into the device being described for the purpose of operational security planning by the healthcare delivery organization. This section supports controls in the RDMP section.

IEC TR 80001-2-2:2012 NIST SP 800-53 Rev. 4 ISO 27002:2013

SBOM-1	Is the SBoM for this product available?	No	This can be made available on request
SBOM-2	Does the SBoM follow a standard or common method in describing software components?	Yes	—
SBOM-2.1	Are the software components identified?	Yes	—
SBOM-2.2	Are the developers/manufacturers of the software components identified?	Yes	—
SBOM-2.3	Are the major version numbers of the software components identified?	Yes	—
SBOM-2.4	Are any additional descriptive elements identified?	N/A	—

SBOM-3	Does the device include a command or process method available to generate a list of software components installed on the device?	Yes	—
SBOM-4	Is there an update process for the SBOM?	Yes	—

SYSTEM AND APPLICATION HARDENING (SAHD)

The device's inherent resistance to cyber attacks and malware.

SAHD-1	Is the device hardened in accordance with any industry standards?	No	—
SAHD-2	Has the device received any cybersecurity certifications?	No	—
SAHD-3	Does the device employ any mechanisms for software integrity checking?	Yes	—
SAHD-3.1	Does the device employ any mechanism (e.g., release-specific hash key, checksums, digital signature, etc.) to ensure the installed software is manufacturer-authorized?	Yes	—
SAHD-3.2	Does the device employ any mechanism (e.g., release-specific hash key, checksums, digital signature, etc.) to ensure the software updates are the manufacturer-authorized updates?	Yes	—
SAHD-4	Can the owner/operator perform software integrity checks (i.e., verify that the system has not been modified or tampered with)?	N/A	—
SAHD-5	Is the system configurable to allow the implementation of file-level, patient level, or other types of access controls?	Yes	ICS is a software product with integrates with Windows Active Directory and access controls can be implemented through Active Directory
SAHD-5.1	Does the device provide role-based access controls?	No	—
SAHD-6	Are any system or user accounts restricted or disabled by the manufacturer at system delivery?	N/A	—
SAHD-6.1	Are any system or user accounts configurable by the end user after initial configuration?	N/A	—
SAHD-6.2	Does this include restricting certain system or user accounts, such as service technicians, to least privileged access?	N/A	—
SAHD-7	Are all shared resources (e.g., file shares) which are not required for the intended use of the device disabled?	No	—
SAHD-8	Are all communication ports and protocols that are not required for the intended use of the device disabled?	Yes	Spacelabs can provide the necessary ports and protocols for customer to configure.
SAHD-9	Are all services (e.g., telnet, file transfer protocol [FTP], internet information server [IIS], etc.), which are not required for the intended use of the device deleted/disabled?	No	—
SAHD-10	Are all applications (COTS applications as well as OS-included applications, e.g., MS Internet Explorer, etc.) which are not required for the intended use of the device deleted/disabled?	No	—
SAHD-11	Can the device prohibit boot from uncontrolled or removable media (i.e., a source other than an internal drive or memory component)?	N/A	—
SAHD-12	Can unauthorized software or hardware be installed on the device without the use of physical tools?	N/A	—
SAHD-13	Does the product documentation include information on operational network security?	N/A	—
SAHD-14	Can the device be hardened beyond the default provided state?	Yes	—
SAHD-14.1	Are instructions available from vendor for increased hardening?	Yes	—
SHAD-15	Can the system prevent access to BIOS or other bootloaders during boot?	N/A	—
SAHD-16	Have additional hardening methods not included in 2.3.19 been used to harden the device?	N/A	—

IEC TR 80001-2-2:2012

NIST SP 800-53 Rev. 4

ISO 27002:2013

		CM-7	A.12.5.1*
Section 5.15, SAHD		AC-17(2)/IA-3	A.6.2.1, A.6.2.2, A.13.1.1, A.13.2.1, A.14.1.2/None
Section 5.15, SAHD		SA-12(10)	A.14.2.7, A.15.1.1, A.15.1.2, A.15.1.3
Section 5.15, SAHD		CM-8	A.8.1.1, A.8.1.2
Section 5.15, SAHD		AC-3	A.6.2.2, A.9.1.2, A.9.4.1, A.9.4.4, A.9.4.5, A.13.1.1, A.14.1.2, A.14.1.3, A.18.1.3
Section 5.15, SAHD		CM-7	A.12.5.1*
Section 5.15, SAHD		CM-7	A.12.5.1*
Section 5.15, SAHD		CM-8	A.8.1.1, A.8.1.2
Section 5.15, SAHD		CM-7	A.12.5.1*
Section 5.15, SAHD		CM-7	A.12.5.1*
Section 5.15, SAHD		CM-7	A.12.5.1*
Section 5.15, SAHD		SA-18	None
Section 5.15, SAHD		CM-6	None
Section 5.15, SAHD		SI-2	A.12.6.1, A.14.2.2, A.14.2.3, A.16.1.3

SECURITY GUIDANCE (SGUD)				IEC TR 80001-2-2:2012	NIST SP 800-53 Rev. 4	ISO 27002:2013
<i>Availability of security guidance for operator and administrator of the device and manufacturer sales and service.</i>						
SGUD-1	Does the device include security documentation for the owner/operator?	Yes	Security manual	Section 5.16, SGUD	AT-2/PL-2	A.7.2.2, A.12.2.1/A.14.1.1
SGUD-2	Does the device have the capability, and provide instructions, for the permanent deletion of data from the device or media?	N/A	—	Section 5.16, SGUD	MP-6	A.8.2.3, A.8.3.1, A.8.3.2, A.11.2.7
SGUD-3	Are all access accounts documented?	N/A	—	Section 5.16, SGUD	AC-6,IA-2	A.9.1.2, A.9.2.3, A.9.4.4, A.9.4.5/A.9.2.1
SGUD-3.1	Can the owner/operator manage password control for all accounts?	N/A	—			
SGUD-4	Does the product include documentation on recommended compensating controls for the device?	N/A	—			
HEALTH DATA STORAGE CONFIDENTIALITY (STCF)				IEC TR 80001-2-2:2012	NIST SP 800-53 Rev. 4	ISO 27002:2013
<i>The ability of the device to ensure unauthorized access does not compromise the integrity and confidentiality of personally identifiable information stored on the device or removable media.</i>						
STCF-1	Can the device encrypt data at rest?	N/A	—	Section 5.17, STCF	SC-28	A.8.2.3
STCF-1.1	Is all data encrypted or otherwise protected?	N/A	—			
STCF-1.2	Is the data encryption capability configured by default?	N/A	—			
STCF-1.3	Are instructions available to the customer to configure encryption?	N/A	—			
STCF-2	Can the encryption keys be changed or configured?	N/A	—	Section 5.17, STCF	SC-28	A.8.2.3
STCF-3	Is the data stored in a database located on the device?	N/A	—			
STCF-4	Is the data stored in a database external to the device?	Yes	—			
TRANSMISSION CONFIDENTIALITY (TXCF)				IEC TR 80001-2-2:2012	NIST SP 800-53 Rev. 4	ISO 27002:2013
<i>The ability of the device to ensure the confidentiality of transmitted personally identifiable information.</i>						
TXCF-1	Can personally identifiable information be transmitted only via a point-to-point dedicated connection?	No	—	Section 5.18, TXCF	CM-7	A.12.5.1
TXCF-2	Is personally identifiable information encrypted prior to transmission via a network or removable media?	No	—	Section 5.18, TXCF	CM-7	A.12.5.1
TXCF-2.1	If data is not encrypted by default, can the customer configure encryption options?	No	—			
TXCF-3	Is personally identifiable information transmission restricted to a fixed list of network destinations?	See Notes	ICS is a software product. It is recommended that customers follow the Spacelabs networking deployment guide.	Section 5.18, TXCF	CM-7	A.12.5.1
TXCF-4	Are connections limited to authenticated systems?	See Notes	ICS is a software product. It is recommended that customers follow the Spacelabs networking deployment guide.	Section 5.18, TXCF	CM-7	A.12.5.1
TXCF-5	Are secure transmission methods supported/implemented (DICOM, HL7, IEE 11073)?	N/A	—			
TRANSMISSION INTEGRITY (TXIG)				IEC TR 80001-2-2:2012	NIST SP 800-53 Rev. 4	ISO 27002:2013
<i>The ability of the device to ensure the integrity of transmitted data.</i>						
TXIG-1	Does the device support any mechanism (e.g., digital signatures) intended to ensure data is not modified during transmission?	No	—	Section 5.19, TXIG	SC-8	A.8.2.3, A.13.1.1, A.13.2.1, A.13.2.3, A.14.1.2, A.14.1.3

TXIG-2	Does the device include multiple sub-components connected by external cables?	N/A	—
--------	---	-----	---

REMOTE SERVICE (RMOT)				IEC TR 80001-2-2:2012	NIST SP 800-53 Rev. 4	ISO 27002:2013
	<i>Remote service refers to all kinds of device maintenance activities performed by a service person via network or other remote connection.</i>					
RMOT-1	Does the device permit remote service connections for device analysis or repair?	No	Product and host server are hosted I customer infrastruture. Customer control can facilitate remote access		AC-17	A.6.2.1, A.6.2.2, A.13.1.1, A.13.2.1, A.14.1.2
RMOT-1.1	Does the device allow the owner/operator to initiative remote service sessions for device analysis or repair?	No	—			
RMOT-1.2	Is there an indicator for an enabled and active remote session?	No	—			
RMOT-1.3	Can patient data be accessed or viewed from the device during the remote session?	No	—		AC-17	A.6.2.1, A.6.2.2, A.13.1.1, A.13.2.1, A.14.1.2
RMOT-2	Does the device permit or use remote service connections for predictive maintenance data?	No	—			
RMOT-3	Does the device have any other remotely accessible functionality (e.g. software updates, remote	N/A	—			

OTHER SECURITY CONSIDERATIONS (OTHR)

NONE

Notes:

Example note. Please keep individual notes to one cell. Please use separate notes for separate information

Note 1

IEC TR 80001-2-2:2012

NIST SP 800-53 Rev. 4

ISO 27002:2013