

## Security Advisory

### SafeNSound Vulnerability Assessment and Potential Product Impact Statement

Ref Doc ID	Version	Release Date	Advisory Status	Related CVE(s)
079-0285-00	A	2025 Sept 4	ACTIVE	N/A

#### 1. VULNERABILITY OVERVIEW

Spacelabs Healthcare has discovered a potential vulnerability that could potentially expose user credentials. Spacelabs has found no evidence of any unauthorized access, but are continuing to assess. This notice is simply to keep you fully informed and reassured of our commitment to transparency and security.

#### 2. RISK ASSESSMENT SUMMARY

Spacelabs internally identified the potential exposure of sensitive information - user credentials on our products. Our assessment has found that Spacelabs SafeNSound could be affected by this vulnerability.

As Spacelabs continue to gain a deeper understanding of the impact of this vulnerability, we will continue to publish technical information to help customers detect, investigate, and mitigate the vulnerability across all our products where applicable.

#### 3. RECOMMENDATIONS

##### GENERAL SECURITY RECOMMENDATIONS

- Monitor and maintain account provisioning and access control based on the principle of least privilege.
- Change passwords on a regular basis.
- Enable multi-factor authentication where possible.

#### 4. EXAMINATION OF SPACELABS PRODUCTS

##### 4.1 ASSESSMENT OF SPACELABS PRODUCTS

In response to the publication of this vulnerability, Spacelabs has conducted an assessment to identify devices potentially at risk of this vulnerability. Please note information is subject to change as the situation evolves.

##### SafeNSound (SNS)

PRODUCT	OPERATING SYSTEM	IMPACT ASSESSMENT
Spacelabs Cloud	Varies	Impacted

PRODUCT	OPERATING SYSTEM	IMPACT ASSESSMENT
SafeNSound	Not applicable	Impacted
SafeNSound desktop	Not applicable	Impacted
SafeNSound mobile	Not Applicable	Impacted

## 5. Document History

Version	Release Date	Purpose
VV00500 Rev A	4 Sept 2025	SafeNSound Credential potential exposure Vulnerability Assessment and Potential Product Impact Statement.

## Disclaimer

The information provided in this advisory is provided "as is" without warranty of any kind. The information in this document is subject to change without notice. Spacelabs disclaims all warranties, either express or implied, including the warranties of merchantability and fitness for a particular purpose. In no event will Spacelabs or any of its suppliers be liable for direct, indirect, special, incidental, or consequential damages of any nature or kind arising from the use of this document, even if Spacelabs or its suppliers have been advised of the possibility of such damages.

This document contains confidential and proprietary language and may not be reproduced or shared with a third party without written permission from Spacelabs. All rights to registrations and trademarks reside with their respective owners.

## Updated copyright to 2025:

©2025 Spacelabs Healthcare. All rights reserved.