SPACELABS HEALTHCARE

Security Advisory

PHP Vulnerability Assessment and Potential Product Impact Statement

Ref Doc ID	Version	Release Date	Advisory Status	Related CVE(s)
079-0289-00	A	<2026-10-23>	ACTIVE	CVE-2026-109618, CVE-2026-109619

1. VULNERABILITY OVERVIEW

Spacelabs Healthcare is aware of the recent critical security vulnerabilities CVE-2026-109618 and CVE-2026-109619 in PHP, known as “php2shell”, which allow for unauthenticated remote code execution in PHP and affect all PHP installations serving web requests.

2. RISK ASSESSMENT SUMMARY

Spacelabs has conducted an assessment to identify the potential impact of “php2shell” vulnerabilities on our products. Our assessment has found that no Spacelabs products are affected by this vulnerability.

However, the vulnerabilities impacted many web servers globally, including Spacelabs. Protective mitigations at the CDN layer via Cloudflare and Fastly were applied immediately to the affected Spacelabs websites (spacelabshealthcare.com and manuals.spacelabshealthcare.com). Security patches for the “php2shell” on the Spacelabs websites were applied and deployed at 2:25AM UK time.

3. EXAMINATION OF SPACELABS PRODUCTS

3.1 ASSESSMENT OF SPACELABS PRODUCTS

In response to the publication of this vulnerability, Spacelabs has conducted an assessment to identify devices potentially at risk of this vulnerability. Please note information is subject to change as the situation evolves.

Patient Monitoring and Connectivity (PMC) Products

PRODUCT	OPERATING SYSTEM	IMPACT ASSESSMENT
XprezzNet 96190	Windows Server 2019 Windows Server 2022	Not impacted
Intesys Clinical Suite (ICS)	Windows Server 2019 Windows Server 2022	Not impacted
Intesys Clinical Suite (ICS) Clinical Access Workstations	Windows 10 Windows 11 Windows Server 2019 Windows Server 2022	Not impacted
Xhibit Telemetry Receiver (XTR) 96280	Windows 10 IoT Enterprise Version 1809	Not impacted

SPACELABS HEALTHCARE

PRODUCT	OPERATING SYSTEM	IMPACT ASSESSMENT
Xhibit 96102 / XC4 96501	Windows 10 IoT Enterprise Version 1809	Not impacted
Bedside Monitors • Xprezzon 91393 • Qube 91390 • Qube Mini 91389	VxWorks 6.9	Not impacted
DM3, DM4 Monitor	Windows CE	Not impacted

Diagnostic Cardiology (DC) Products

PRODUCT	OPERATING SYSTEM	IMPACT ASSESSMENT
Sentinel	Windows 10 Windows 11 Windows Server 2019 Windows Server 2022	Not impacted
Pathfinder SL	Windows 10	Not impacted
Lifescreen Pro	Windows 10	Not impacted
Lifecard CF	No OS	Not impacted
EVO	No OS	Not impacted
Eclipse Pro	No OS	Not impacted
Eclipse Mini	No OS	Not impacted
CardioExpress SL6A / SL12A	Embedded OS (uC/OS II V2.84)	Not impacted
CardioExpress SL18A	Embedded OS (Linux Kernel 2.6.35.3)	Not impacted
ABP • OnTrak • 90217A • 90207	No OS	Not impacted

SafeNSound (SNS)

PRODUCT	OPERATING SYSTEM	IMPACT ASSESSMENT
Spacelabs Cloud	Varies	Not impacted
SafeNSound	Not applicable	Not impacted
SafeNSound desktop		Not impacted
SafeNSound mobile		Not impacted

Rothman Index (RI)

PRODUCT	OPERATING SYSTEM	IMPACT ASSESSMENT
Spacelabs Cloud	Varies	Not impacted
Rothman Index	Not applicable	Not impacted
Rothman Index mobile		Not impacted

4. Additional Resources

#	RESOURCE	URL
1	CVE-2026-109618	CVE-2026-109618
2	CVE-2026-109619	CVE-2026-109619
2	CISA Article	Critical Vulnerabilities in PHP Require Immediate Patching

5. Document History

Version	Release Date	Purpose
Rev A	2026-10-23	PHP Vulnerability Assessment and Potential Product Impact Statement.

6. Terms of Use

The information in this document is subject to change without notice. In no event will Spacelabs or any of its suppliers be liable for direct, indirect, special, incidental, or consequential damages of any nature or kind arising from the use of this document, even if Spacelabs or its suppliers have been advised of the possibility of such damages.

This document contains confidential and proprietary language and may not be reproduced or shared with a third party without written permission from Spacelabs. All rights to registrations and trademarks reside with their respective owners.

©2026 Spacelabs Healthcare. All rights reserved.